

Title (en)
HEARING AID FOR INDICATING A PATHOLOGICAL CONDITION

Title (de)
HÖRGERÄT ZUR ANZEIGE EINES PATHOLOGISCHEN ZUSTANDS

Title (fr)
PROTHÈSE AUDITIVE POUR INDIQUER UN ÉTAT PATHOLOGIQUE

Publication
EP 3139638 A1 20170308 (EN)

Application
EP 15184032 A 20150907

Priority
EP 15184032 A 20150907

Abstract (en)
According to an embodiment, a hearing aid adapted to indicate a pathological condition of a user of the hearing aid when the hearing aid is in use is disclosed. The hearing aid includes an input transducer, a processing unit, an output transducer and a feedback measurement unit. The input transducer is adapted to transform an input sound signal into an electrical input signal. The processing unit adapted to generate an electrical output signal by processing the electrical input signal. The output transducer is adapted to transform the processed electrical output signal to an acoustic output signal and to direct the acoustic output signal towards a tympanic membrane of the user. The feedback measurement unit is adapted to measure an acoustic feedback produced by the tympanic membrane in response to the acoustic output signal. During a time duration, a feedback path logger is adapted to continuously track the measured acoustic feedback and a feedback path modeller is adapted to generate a base pattern across a plurality of frequencies of the tracked acoustic feedback. During a subsequent time duration, the feedback path logger is adapted to continuously track the measured acoustic feedback and the feedback path modeller is adapted to generate a measured pattern across the plurality of frequencies of the tracked acoustic feedback. The hearing aid further includes a comparator and a state classifier. The comparator is adapted to determine a variation between the measured pattern and the base pattern and the state classifier adapted to identify at least one pathological condition corresponding to the determined variation.

IPC 8 full level
H04R 25/02 (2006.01); **A61B 5/12** (2006.01)

CPC (source: EP)
H04R 25/453 (2013.01); **H04R 2225/39** (2013.01)

Citation (applicant)

- US 2015063614 A1 20150305 - PETERSEN SVEND OSCAR [DK], et al
- US 2015043764 A1 20150212 - RASMUSSEN KARSTEN BO [DK]
- US 2014146977 A1 20140529 - MUNK STEEN MICHAEL [DK], et al
- US 2013188796 A1 20130725 - KRISTENSEN MICHAEL SMED [DK], et al
- US 2013070936 A1 20130321 - JENSEN JESPER [DK], et al
- GUO MENG, ANALYSIS, DESIGN, AND EVALUATION OF ACOUSTIC FEEDBACK CANCELLATION SYSTEMS FOR HEARING AIDS, 2013
- A. SPRIET ET AL.: "Adaptive feedback cancellation in hearing aids", JOURNAL OF THE FRANKLIN INSTITUTE, vol. 343, 2006, pages 545 - 573

Citation (search report)

- [XAI] WO 2009023633 A1 20090219 - PERSONICS HOLDINGS INC [US], et al
- [A] WO 2012082721 A2 20120621 - UNIV ILLINOIS [US], et al
- [A] WO 2013102867 A1 20130711 - CONSIGLIO NAZIONALE RICERCHE [IT]
- [A] MARGOLIS ROBERT H ET AL: "Wideband reflectance tympanometry in normal adults", THE JOURNAL OF THE ACOUSTICAL SOCIETY OF AMERICA, AMERICAN INSTITUTE OF PHYSICS FOR THE ACOUSTICAL SOCIETY OF AMERICA, NEW YORK, NY, US, vol. 106, no. 1, 1 July 1999 (1999-07-01), pages 265 - 280, XP012001058, ISSN: 0001-4966, DOI: 10.1121/1.427055

Cited by
WO2018154143A1; CN113260305A; EP3675524A1; EP3675526A1; US11317222B2; US11736868B2

Designated contracting state (EPC)
AL AT BE BG CH CY CZ DE DK EE ES FI FR GB GR HR HU IE IS IT LI LT LU LV MC MK MT NL NO PL PT RO RS SE SI SK SM TR

Designated extension state (EPC)
BA ME

DOCDB simple family (publication)
EP 3139638 A1 20170308

DOCDB simple family (application)
EP 15184032 A 20150907