



(19)

Europäisches Patentamt

European Patent Office

Office européen des brevets



(11)

EP 0 786 796 A1

(12)

EUROPEAN PATENT APPLICATION

(43) Date of publication:
30.07.1997 Bulletin 1997/31

(51) Int. Cl.⁶: H01J 49/42

(21) Application number: 97104015.9

(22) Date of filing: 28.05.1993

(84) Designated Contracting States:
CH DE FR GB IT LI

(30) Priority: 29.05.1992 US 890991

(62) Document number(s) of the earlier application(s) in
accordance with Art. 76 EPC:
93108670.6 / 0 575 777

(71) Applicant: VARIAN ASSOCIATES, INC.
Palo Alto, California 94304-1030 (US)

(72) Inventors:
• Wells, Gregory J.
Fairfield, California 94533 (US)

• Wang, Mingda
Walnut Creek, California 94598 (US)

(74) Representative: Kahler, Kurt, Dipl.-Ing. et al
Patentanwälte
Kahler, Käck, Fiener et col.,
Vorderer Anger 268
86899 Landsberg/Lech (DE)

Remarks:

This application was filed on 11 - 03 - 1997 as a
divisional application to the application mentioned
under INID code 62.

(54) Methods of using ion trap mass spectrometers

(57) In a CI process it is desirable to optimize the number of product ions that undergo mass analysis. If there are too few product ions, the mass analysis will be noisy, and if there are too many product ions resolution and linearity will be lost. The formation of product ions is a function of the number of reagent ions present in the trap, the number of sample molecules in the trap, the reaction rate between the reagent ions and the sample ions, and the reaction time during which reagent ions are allowed to react with sample molecules. Thus, the application covers a two-step process consisting of first conducting a "prescan" of the contents of the ion trap to obtain a gross determination of the number of product ions in the trap, followed by a mass analysis scan with the parameters of mass analysis scan being adjusted based on the data collected during the prescan.




FIG. 2

EP 0 786 796 A1

Description

Field of the Invention

The present invention relates to methods of using ion trap mass spectrometers ("ion traps") by applying supplemental voltages to the trap, and is particularly related to methods of operating ion traps in the chemical ionization mode, and for conducting multiple mass spectroscopy experiments ("MS").

Background of the Invention

The quadrupole ion trap, sometimes referred to as an ion store or an ion trap detector, is a well-known device for performing mass spectroscopy. A ion trap comprises a ring electrode and two coaxial end cap electrodes defining an inner trapping volume. Each of the electrodes preferably has a hyperbolic surface, so that when appropriate AC and DC voltages (conventionally designated "V" and "U", respectively) are placed on the electrodes, a quadrupole trapping field is created. This may be simply done by applying a fixed frequency (conventionally designated "f") AC voltage between the ring electrode and the end caps. The use of an additional DC voltage is optional.

Typically, an ion trap is operated by introducing sample molecules into the ion trap where they are ionized. Depending on the operative trapping parameters, ions may be stably contained within the trap for relatively long periods of time. Under certain trapping conditions, a large range of masses may be simultaneously held within the trap. Various means are known for detecting ions that have been so trapped. One known method is to scan one or more of the trapping parameters so that ions become sequentially unstable and leave the trap where they may be detected using an electron multiplier or equivalent detector. Another method is to use a resonance ejection technique whereby ions of consecutive masses can be sequentially scanned out of the trap and detected.

The mathematics of the trapping field, although complex, are well developed. Ion trap users are generally familiar with the stability envelop diagram depicted in FIG. 1. For a trap of a given radius r_0 and for given values of U, V and f, whether an ion of mass-to-charge ratio (m/e) will be trapped depends on the solution to the following two equations:

$$a_z = \frac{-8eU}{m r_0^2 \omega^2}$$

$$q_z = \frac{4eV}{m r_0^2 \omega^2}$$

Where ω is equal to $2\pi f$.

Solving these equations yields values of a and q for

a given m/e. If, for a given ion, the point (a,q) is inside the stability envelop of FIG. 1, the ion will be trapped by the quadrupole field. If the point (a,q) falls outside the stability envelop, the ion will not be trapped and any such ions that are created within the trap will quickly depart. It follows that by changing the values of U, V or f one can control whether a particular mass ion is trapped in the quadrupole field. It should be noted that it is common in the field to use the terms mass and mass-to-charge ratio interchangeably. However, strictly speaking, it proper to use the term mass-to-charge ratio.

In the absence of a DC voltage, the equations set forth actually relate to stability in the direction of the z axis, i.e., the direction of the axis of the electrodes. Ions will become unstable in this direction before becoming unstable in the r direction, i.e., a direction radial to the axis. Thus, it is normal to limit consideration of stability to z direction stability. The differential in stability results in the fact that unstable ions will leave the trap in the z direction, i.e., axially.

In commercially available implementations of the ion trap, the DC voltage, U, is set at 0. As can be seen from the first of the above equations, when $U = 0$, then $a_z = 0$ for all mass values. As can be seen from the second of the above equations the value of q_z will be inversely proportional to the mass of the particle, i.e., the larger the value of the mass the lower the value of q_z . Likewise, the higher the value of V the higher the value of q_z . Turning to the FIG. 1 stability envelop, it can also be seen that for the case where $U = 0$, and for a given value of V, all masses above a certain cut-off value will be trapped in the quadrupole field. Although all masses above a cut-off value are stable in such a trapping field, there are limits to the quantity of ions of a particular mass value that will be trapped due to space charge effects. As discussed below such quantity limitations are also a function of the magnitude of V.

Several methods are known for ionizing sample molecules within the ion trap. Perhaps the most common method is to expose the sample to an electron beam. The impact of electrons with the sample molecules cause them to become ionized. This method is commonly referred to as electron impact ionization or "EI".

Another commonly used method of ionizing sample with an ion trap is chemical ionization or "CI". Chemical ionization involves the use of a reagent gas which is ionized, usually by EI within the trap, and allowed to react with sample molecules to form sample ions. Commonly used reagent gases include methane, isobutane, and ammonia. Chemical ionization is considered to be a "softer" ionization technique. With many samples CI produces fewer ion fragments than the EI technique, thereby simplifying mass analysis. Chemical ionization is a well known technique that is routinely used not only with quadrupole ion traps, but also with most other conventional types of mass spectrometers such as quadrupole mass filters, etc.

Other, more specialized, methods of ionization are also in use in mass spectroscopy. For example, photoionization is a well known technique that, similar to electron impact ionization, will affect all molecules contained in the trap.

Most ion trap mass spectrometer systems in use today include a gas chromatograph ("GC") as a sample separation and introduction device. When using a GC for this purpose, sample which elutes from the GC continuously flows into the mass spectrometer, which is set up to perform periodic mass analyses. Such analyses may, typically, be performed at a frequency of about one scan per second. This frequency is acceptable since peaks typically elute from a modern high resolution GC over a period of several seconds to many tens of seconds. When performing CI experiments in such a system, a continuous flow of reagent gas is maintained. As a practical matter it is undesirable to interrupt the flow of sample gas from the GC to the ion trap. Likewise, when conducting both CI and EI experiments on a sample stream, it is undesirable to interrupt the flow of reagent gas to the ion trap.

When performing CI, it is necessary to ionize a reagent gas, which then chemically reacts with and ionizes the sample gas. As noted, electron impact ionization within the ion trap is the preferred method of ionizing the reagent gas. However, if sample is present in the ion trap when the electron beam is turned on to ionize the reagent gas, the sample will also be subject to EI. As noted above, where chromatography is used to separate a sample before it is introduced into the ion trap, it is impractical to interrupt the flow of sample gas. Therefore, there is not a practical way to ionize the reagent gas without also ionizing the sample. Thus, unless mitigating measures are taken, sample ions will be formed by both CI and EI, leading to potentially confused results.

The prior art solution to this problem is described in U.S. Pat. No. 4,686,367, entitled *Method of Operating Quadrupole Ion Trap Chemical Ionization Mass Spectrometer*, issued on August 11, 1987, to Louris, *et al.* The method of the '367 patent seeks to minimize the effects of EI of the sample by minimizing the number of sample ions trapped by the ion trap while reagent gas is being ionized. The method that is taught for doing this is to apply a low value of V to the trap during the EI step so that the low mass reagent ions will be trapped, but the number of high mass ions will be small. In the words of the patent, "at sufficiently low RF values, [*i.e.*, values of V] high molecular weight ions are not efficiently trapped. So, at low RF voltages only the low mass ions are stored." (Column 5, lines 33 - 36.)

As is explained above, when operating using the RF only method, which is preferred in the '367 patent and which is the method used in all known commercial embodiments of the ion trap, the trap inherently traps all masses above a cut-off mass which is set by the value of the RF trapping voltage. Thus, to trap low mass ions, whether they be reagent ions or sample ions, it is nec-

essary to set V at a sufficiently low value. When V is set low enough the trap inherently has a poor efficiency in trapping high mass ions due to space charge effects. A theoretical way of looking at this is that the volume of the interior of the ion trap which stores ions of a particular mass is proportional to the value of V and is inversely proportionally to the mass. Thus, for any given V a smaller volume of the ion trap is available to store high mass ions than low mass ones. When the volume is quite small the number of ions than can be stored is reduced due to space charge effects.

It should be noted that setting a low value of V does not cause all high mass ions to leave the trap; such ions continue to have values of *a* and *q* that map into the stability envelop. All that can be done following the technique of the '367 patent is to reduce the number of high mass ions in the trap during the EI step. In this respect, the statement in the patent that "at low RF voltages only the low mass ions are stored" appears to be incorrect. As described below, experimental results show the presence of detectable quantities of high mass ions created by EI in experiments conducted using the method of the '367 patent. Moreover, the number of high mass ions that remain trapped will depend on the mass, so that a substantial number of sample ions close, yet higher, in mass than the reagent ions, will be trapped.

Some reagent molecules form a variety of ions having different masses. Ionization at RF voltages substantially below that necessary to trap the lowest mass reagent ion, which is necessary to remove most of the high mass sample ions, will reduce the number of reagent ions that are trapped, as well as the high mass sample ions. This effect is related to mass so that the higher mass reagent ions will be disproportionately lost from the trap.

A related problem exists when conducting both EI and CI experiments on a single sample stream in an ion trap. As noted above, for practical reasons it is undesirable to stop the flow of reagent gas to the trap. However, if reagent gas is present when an EI experiment is run, the reagent gas will be ionized creating reagent gas ions which may cause CI of the sample unless they are eliminated from the trap before reactions can occur. This problem does not exist when conducting only EI experiments on a sample stream since the reagent gas flow may simply be kept off during such experiments.

The method of the lowering the trapping voltage is not applicable, however, to solving this problem since it would not eliminate low mass reagent ions from the trap. One solution used to solve this problem, as taught in the '367 patent, is to raise the RF trapping voltage so as not to store the low mass reagent ions. However, this has the undesired effect of changing the trapping conditions from those which are normally used. For example, when the trapping voltage is set to store ions of mass 20 and above, the average ionizing energy of electron entering the trap is 70 eV. Raising the trapping voltage to store only ions of mass 45 and above, so as to eliminate methane reagent ions at mass 43, would double

the average electron energy. Such an increase would change the mass spectrum of many compounds and would reduce the trapping efficiency for the sample ions.

In a CI process it is desirable to optimize the number of product ions that undergo mass analysis. If there are too few product ions, the mass analysis will be noisy, and if there are too many product ions resolution and linearity will be lost. The formation of product ions is a function of the number of reagent ions present in the trap, the number of sample molecules in the trap, the reaction rate between the reagent ions and the sample ions, and the reaction time during which reagent ions are allowed to react with sample molecules. One can increase the number of reagent ions present in the trap by increasing the EI ionization time, *i.e.*, keeping the electron beam on a longer time. Likewise, one can increase the number of sample ions formed in the trap by increasing the reaction time.

One prior art method of addressing this issue is set forth in U.S. Pat. No. 4,771,172, entitled *Method Of Increasing The Dynamic Range And Sensitivity Of A Quadrupole Ion Trap Mass Spectrometer Operating In The Chemical Ionization Mode*, issued on September 13, 1988, to Weber-Grabau, *et al.* This patent covers a method of adjusting the parameters used in an ion trap in the CI mode so as to optimize the results. In order to optimize the parameters, the patent teaches the method of performing a CI "prescan," done in accordance with the method of the '367 patent, preceding each mass analysis. This prescan is a complete CI scan cycle in which the ionization and reaction times are fixed at values smaller than those that would be used in a normal analytical scan, and in which the product ions are scanned out of the trap faster than in a normal analytical scan. The resulting product ions that are ejected from the trap during the prescan are not mass resolved and the ion signal is only integrated to give a total product ion signal. During the prescan the total number of product ions in the trap are measured and the parameters, *i.e.*, the ionization time and/or the reaction time for the subsequent mass analysis scan are adjusted.

Thus, the patent covers a two-step process consisting of first conducting a "prescan" of the contents of the ion trap to obtain a gross determination of the number of product ions in the trap, followed by a mass analysis scan of the type taught in the '367 patent, with the parameters of mass analysis scan being adjusted based on the data collected during the prescan. The disadvantage of the prior art method of extending the dynamic range by using a prescan to estimate the sample amounts in the trap is that it requires additional time to perform the prescan, and thus fewer analytical scans can be performed in the same time period. Not only does each of the prescans consume time, but each produces data which has no independent value apart from its use in adjusting the parameters for the mass analysis scan. However, adjustments in the mass analysis scan parameters are only required when conditions change.

It is not necessary to make adjustments for each scan and, thus, in many instances the prescan step, in addition to consuming time, will not serve any useful purpose. Thus, there is a need for an improved method of adjusting the ion trap during chemical ionization experiments to operate within its dynamic range.

There is a demand to employ the ion trap mass spectrometer in conducting so-called MSⁿ experiments. In MSⁿ experiments, a single ion species is isolated in the trap and is dissociated into fragments. The fragments created directly from the sample species are known in the art as daughter ions, and the sample is referred to as the parent ion. The daughter ions may also be fragmented to create granddaughter ions, etc. The value of n refers to the number of ion generations that are formed; thus, in an MS² or MS/MS experiment, only daughter ions are formed and analyzed.

A prior art method of conducting MSⁿ experiments is described in U.S. Pat. No. 4,736,101, entitled *Method Of Operating Ion Trap In MS/MS Mode*, issued April 5, 1988 to Syka, *et al.* After isolating an ion species of interest, the parent ions are resonantly excited by means of a single supplemental AC frequency which is tuned to the resonant frequency of the ions of interest. The amplitude of the supplemental frequency is set at a level which causes the ions to gain energy so that their oscillations within the trap are greater, but which is not large enough to cause the ions to be ejected from the trap. As the ions oscillate within the trap they collide with molecules of the damping gas in the trap and undergo collision induced dissociation thereby forming daughter ions. By applying resonant frequencies associated with the mass-to-charge ratios of the daughter ions, they can similarly be fragmented.

The difficulty with the method of the '101 patent is that the precise resonant frequency of the ions of interest cannot be determined *a priori* but must be determined *a posteriori*. The resonant frequency of an ion, also referred to as its secular frequency, varies with the ion mass-to-charge ratio, the number of ions in the trap, hardware variances and other parameters which cannot be precisely determined in a simple way. Thus, the precise resonant frequency must of an ion species be determined empirically. While empirical determination can be performed without great difficulty when a static sample is introduced into the trap, it is quite difficult to accomplish when a dynamic sample, such as the output of a GC, is used.

One prior art approach to overcoming the foregoing problem in determining the precise resonant frequency of a sample ion of interest is to use a broadband excitation centered around the calculated frequency. For example, such a broadband excitation may have a bandwidth of about 10 KHz. Another method is to conduct a frequency prescan, *i.e.*, sweep the supplemental field across a frequency range in the area of interest and observe the resonant frequency empirically. However, neither of these solutions are particularly satisfactory.

Accordingly, it is the object of the present invention relates to a method of optimizing the experimental parameters utilized in an ion trap in order to operate within dynamic range of the trap.

Summary of the Invention

The object of the invention is solved by the method as defined in claim 1.

In an embodiment of the present invention, mass spectral data associated with the largest peak measured during one scan of the ion trap is used to adjust, if necessary, experimental parameters utilized during the subsequent scan so that the trap is operated within its dynamic range.

Brief Description of the Drawings

FIG. 1 is a plot of the stability diagram associated with an ion trap.

FIG. 2 is a partially schematic view of apparatus used to practice the method of the present inventions.

FIG. 3 is a graph showing the control of the supplemental broadband AC field in relation to the gating of the electron beam used for electron impact ionization in accordance with the present invention.

FIGS. 4A - 4G are mass spectra of various samples comparing the present invention with the method of the prior art.

FIG. 5 shows an alternate arrangement of the apparatus of FIG. 2 for use in practicing the present invention.

FIGS. 6A - 6E are mass spectra of various samples showing how the application of a supplemental low frequency field may be used to cause fragmentation of a parent ion within an ion trap.

FIGS. 7A - 7C are mass spectra showing how the application of a supplemental low frequency field may be used to eliminate high mass ions from an ion trap.

FIGS. 8A - 8C are mass spectra showing how the application of a supplemental low frequency field may be used in conducting chemical ionization experiments.

Detailed Description

An apparatus for practicing the present invention is schematically shown in FIG. 2. Ion trap 10, shown schematically in cross-section, comprises a ring electrode 20 coaxially aligned with upper and lower end cap electrodes 30 and 35, respectively. Preferably, the trap electrodes have hyperbolic inner surfaces, although other shapes, for example, electrodes having a cross-sections forming an arc of a circle, may also be used to create trapping fields. The design and construction of ion trap mass spectrometers is well-known to those skilled in the art and need not be described in detail. A commercial model ion trap of the type described herein is sold by the assignee hereof under the model designation Saturn.

Sample gas, for example from a gas chromatograph 40, is introduced into the ion trap 10. Since GC's typically operate at atmospheric pressure while ion traps operate at greatly reduced pressures, pressure reducing means (not shown) are required. Such pressure reducing means are conventional and well known to those skilled in the art. While the present invention is described using a GC as a sample source, the source of the sample is not considered a part of the invention and there is no intent to limit the invention to use with gas chromatographs. Other sample sources, such as, for example, liquid chromatographs with specialized interfaces, may also be used.

Also connected to the ion trap is a source of reagent gas 50 for conducting chemical ionization experiments. Sample and reagent gas that is introduced into the interior of ion trap 10 may be ionized by electron bombardment as follows. A beam of electrons, such as from a thermionic filament 60 powered by filament power supply 65, is controlled by a gate electrode 70. The center of upper end cap electrode 30 is perforated (not shown) to allow the electron beam generated by filament 60 and gate electrode 70 to enter the interior of the trap. The electron beam collides with sample and reagent molecules within the trap thereby ionizing them. Electron impact ionization of sample and reagent gases is also a well-known process that need not be described in greater detail.

A trapping field is created by the application of an AC voltage having a desired frequency and amplitude to stably trap ions within a desired range of mass-to-charge ratios. RF generator 80 is used to create this field, and is applied to the ring electrode. While it is well known that one may also apply a DC voltage to modify the trapping field and to work at a different portion of the stability diagram of FIG. 1, as a practical matter, commercially available ion traps all operate using an AC trapping field only.

A variety of methods are known for determining the mass-to-charge ratios of the ions which are trapped in the ion trap to thereby obtain a mass spectrum of the sample. One known method is to scan the trap so that ions of sequential mass-to-charge ratio are ejected in order. A first known method of scanning the trap is to scan one of the trapping parameters, such as the magnitude of the AC voltage, so that ions sequentially become unstable and leave the trap where they are detected using, for example, electron multiplier means 90.

Another known method of scanning the trap involves use of a supplemental AC dipole voltage applied across end caps 30 and 35 of ion trap 10. Such a voltage may be created by a supplemental waveform generator 100, coupled to the end caps electrodes by transformer 110. The supplemental AC field is used to resonantly eject ions in the trap. Each ion in the trap has a resonant frequency which is a function of its mass-to-charge ratio and of the trapping field parameters. When an ion is excited by a supplemental RF field at its reso-

nant frequency it gains energy from the field and, if sufficient energy is coupled to the ion, its oscillations exceed the bounds of the trap, *i.e.*, it is ejected from the trap. Ions ejected in this manner can also be detected by electron multiplier 90 or an equivalent detector. When using the resonant ejection scanning technique, the contents of the trap can be scanned in sequential order by either scanning the frequency of the supplemental RF field or by scanning one of the trapping parameters such as the magnitude of V, the AC trapping voltage. As a practical matter, scanning the magnitude of the AC voltage is preferred.

In addition, a new method of scanning the ion trap is described hereinbelow.

In one embodiment of the present invention, supplemental RF generator 100, which may also be used for scanning the trap as described above, is capable of generating a broadband RF field which is used to resonantly eject sample ions created by EI during the time that the reagent gas is being ionized. FIG. 3(a) shows the gating of the electron beam used to ionize the reagent gas. Beginning at t1 and ending at t2, electron gate 70 is turned on to allow the electron beam to enter the trap to form reagent ions from the neutral reagent gas. As shown in FIG. 3(b) coincident with the electron gate admitting electrons into the trap, supplemental waveform generator 100 applies a broadband signal to the end caps of the trap, 30, 35, for a period of time that begins at t1 and ends at t3. As shown, the broadband excitation exceeds the gate time. Alternately, the supplemental broadband signal could be applied starting at a time later than t1, or even later than t2, *i.e.*, after the electron ionization is complete. Likewise, the supplemental signal could also start at a time prior to t1. The important aspect being that the supplemental field for elimination of unwanted sample ions be kept "on" for a period of time extending after the end of the period during which ions are created.

The broadband AC voltage applied to the end caps can either be out of phase (dipole excitation) or in phase (quadrupole excitation). An alternative method of obtaining quadrupole excitation is the application of the supplemental waveform to the ring electrode as shown in FIG. 5, rather than to the end caps.

The supplemental waveform contains a range of frequencies of sufficient amplitude to eject unwanted sample ions of mass greater than the highest mass reagent ion, by means of resonant power absorption by the trapped ions. Each of the sample ions is in resonance with a frequency component of the supplementary waveform. Accordingly, they absorb power from the supplementary field and leave the trapping field. After the supplemental field has ejected the unwanted ions it is turned off and the CI reagent ions react with the sample molecules to produce CI sample ions. These ions are then scanned from the trap for detection in a conventional manner as described above.

The supplemental waveform described above is broadband and has a first frequency component corre-

sponding to the lowest mass to be ejected and a last frequency corresponding to the highest mass to be ejected. Between the first and last frequencies are a series of discrete frequency components which may be spaced evenly or unevenly, and which may have phases that are either random or with a fixed functional relationship. The amplitudes of the frequency components can either be uniform or they can be tailored to a functional form so as to compensate for frequency dependencies of the hardware or to compensate for the distribution of q values due to the distribution of the masses that are stored in the trap. The broadband waveform has a sufficient number of frequency components so that any ion with a resonant frequency between the first and last components of the waveform will be resonantly ejected by this supplemental field. Thus, all sample ions formed during EI will be eliminated from the trap before the mass analysis scan and there will be no gaps in the mass range that is affected.

As a practical matter, the reagent gases that are used in CI experiments are all low in molecular weight such that the reagent ions formed during EI of the contents of the trap will, in almost all cases, be lower in mass-to-charge ratio than the sample ions. In the rare instance when a sample ion is created that is lower in mass than the reagent ions, a specific frequency may be added to the broadband excitation to cause that specific mass to be ejected along with others.

The advantage of the invention over prior art is the ability to remove unwanted sample ions formed by EI during the ionization of the CI reagent gas. The ability to reject these ions will allow longer ionization times and greater emission currents to be used, thus increasing the sensitivity of CI.

FIG. 4A shows the residual EI spectrum of a sample of tetrachloroethane using the scan conditions that are used in the prior art method. FIG. 4B shows the elimination of the sample ions formed during the ionization step using the broadband waveform. FIG. 4C shows the residual EI spectrum of a sample of trichloroethane and PFTBA with methane reagent gas present in the trap using the prior art method. FIG. 4D shows the elimination of the sample ions formed during the ionization step using the broadband waveform of the present invention. It can be seen that the reagent ions at mass 43 are still present even though the sample ions that are just above them in mass are removed. FIG. 4E shows the spectrum under the same conditions as in FIG. 4D except that the supplemental waveform is off. FIG. 4F shows a spectrum of hexachlorobenzene using the prior art method. A mixture of EI ion fragments are observed at mass 282, 284, 286, 288 and 290. In addition, ions due to the protonated sample (from CI) are observed at mass 283, 285, 287, 289 and 291. FIG. 4G shows the spectrum using the method described herein. It can be seen that the unwanted ions from the EI process are almost completely removed.

In another aspect of the present invention, data obtain from one scan are used, if necessary, to adjust

the parameters of the subsequent scan to ensure that the trap is operated within its dynamic range. Preferably, the amplitude of the most intense ion of a scan (the base peak) is used to adjust the ionization and/or reaction time for the next scan. The magnitude of the base peak is used to adjust the ionization and reaction times for the subsequent scan so as to maintain a substantially constant number of ions of the base peak. Since most of the charge ejected from the trap during the scan is due to the base peak, it is a good representation of the total amount of charge from the sample in the trap. By keeping the total sample charge nearly constant in the trap the dynamic range of the sample can be increased. Alternately, with the mass spectral information from one scan it is possible to adjust the parameters of the subsequent mass analysis scan to focus, for example, on only particular sample ions of interest, *i.e.*, to optimize for a particular species.

Preferably, when adjusting the parameters for a scan based on the previous scan, both the reaction time and the ionization time are changed in a set ratio. This makes it easier to normalize the results from one scan to the next.

An advantage of this inventive method is the reduction in the scan time for large dynamic range samples. This is accomplished by using the intensity of the base peak from the previous scan as a measure of the amount of sample in the trap; thus eliminating the need for a time-consuming prescan as is used in the prior art.

A broadband supplemental field can also be used to eliminate reagent ions from the trap when conducting an EI experiment. In some instances, the user of an ion trap may wish to conduct both EI and CI experiments on the same sample stream. Under such circumstances, it is undesirable to stop the flow of reagent gas into the trap while conducting EI, yet the presence of reagent ions is likely to cause confused analytic data. By using a supplemental RF broadband excitation, any reagent ions formed during electron impact ionization of the sample can be resonantly ejected from the trap as soon as they are created. The same timing sequence shown in FIG. 3 can be used to practice this aspect of the invention. In this embodiment of the invention, the broadband RF excitation may be constructed in accordance with any of the above-described alternatives, except that the frequency range should be tailored to eliminate only the low mass reagent ions.

Waveform generator 100 of FIG. 2 can also be used to apply a low frequency non-resonant field to perform CI experiments, to conduct MSⁿ experiments and to scan the contents of the trap to obtain a mass spectrum. A low frequency supplemental voltage from waveform generator 100 is applied as a dipole field across end caps 30, 35 of ion trap 10. The frequency of the dipole field is unrelated to the resonant frequencies of any of the ions (whether sample or reagent ions) stored in the trap. The waveform shape is preferably a square wave, but may be almost any shape including sine, sawtooth, triangular waveforms. As noted, the frequency of the

supplemental voltage is relatively low, such as between 100 Hz and several thousand Hz. Experiments suggest that the present invention would work at frequencies below about 10,000 Hz, which is about the beginning of the range of resonant frequencies of sample ions. Preferably, however, the frequency should be in the range of hundreds of Hz.

It is believed that the supplemental squarewave dipole field alternately displaces the center of the pseudo-potential well of the trapping field to different locations along the z-axis. Each time the center of the pseudo-potential well of the trapping field is displaced, trapped ions pick up translational energy from the trapping field and begin to oscillate around the new center. Thus, displacement of the center of the oscillations tends to increase the magnitude of the oscillations. Gradually, as the ions lose energy to the background gas, they move towards the new center. If the center of the pseudo-potential field is again moved, such as when the squarewave changes polarity, the process repeats itself. It can be seen that the frequency of the supplemental dipole field should be low so that ions are able to migrate towards the new center before the field is changed.

When the center of the pseudo-potential well is moved, as described above, the ions begin oscillating about a new point in space becoming more energetic. The energy added to ions will be sufficient to cause many of them to dissociate due to collisions with the damping gas, thereby forming daughter ions. As the process is repeated, more and more of the ions will dissociate in this manner. Another advantage of this method is that it imparts more energy to the ions than resonance excitation and, thus, in some cases, can result in more extensive ion fragmentation.

Since the method described above does not rely on the resonant frequency of the ions in the ion trap, it operates on all ions in the trap simultaneously. Thus, using this method it is possible to simultaneously create several generations of ion fragments without the need to apply resonant frequencies associated with each of the fragments. If desired, prior to practicing the present invention, an ion species of interest could first be isolated in the trap in accordance with known prior art methods.

Using this method it is possible to obtain a complete "fingerprint" of a compound, facilitating the identification of the compound. Mass-to-charge ratio cannot, alone, be used to unambiguously identify a parent ion. However, knowing not only the mass-to-charge ratio of the parent ion, but also the masses of all of the ion fragments can be used to unambiguously identify the parent.

It has also been discovered that applying a low frequency voltage to the ion trap can be used as a mechanism to cause ions having masses above a certain cutoff mass to be eliminated from the ion trap. The cutoff mass is a function of the magnitude of the supplemental low-frequency voltage. One model of how an ion

trap operates is that the ions are, in essence, trapped in a potential well, with the "depth" of the well being a function of, among other things, the mass-to-charge ratio. The higher the mass, the shallower the well. It is believed that the observed phenomenon of elimination of high mass ions by application of a low frequency supplemental field is related to the relatively shallow depth of the potential well associated with high mass ions. In particular, it is believed that the shifting of the center of the pseudo-potential well causes high mass ions to gain sufficient energy to overcome the well barrier and leave the ion trap.

This phenomenon can be used to advantage both in chemical ionization experiments and in scanning the ion trap. As described above, when conducting chemical ionization experiments, it is necessary to eliminate high mass sample ions that are created during EI of the reagent gas. An alternate method of eliminating the sample ions is to apply a low-frequency supplemental field, as described above, having a magnitude which is sufficient to eliminate all sample ions from the trap, while leaving the reagent ions unaffected. The timing sequence for applying this supplemental low-frequency field may be as depicted in FIG. 3, or any of the alternatives timing sequences described above in connection therewith. In this regard, it is noted that the ionization period of FIG. 3(a) which may be less than a millisecond in duration, may be shorter in duration than a half-cycle of the low-frequency supplemental voltage. Thus, the duration of application of the supplemental voltage, as shown in FIG. 3(b), may be much longer in duration, and FIG. 3 is not drawn to scale.

The application of a low-frequency supplemental voltage can also be used as a mechanism for scanning the ion trap to obtain a mass spectrum. This can be done by scanning the magnitude of the supplemental low-frequency voltage. If the supplemental voltage is initially low and is ramped-up, masses will be ejected from the trap sequentially in descending order. Alternately, the low-frequency supplemental voltage can be held constant and one of the trapping parameters scanned to obtain the equivalent effect.

FIG. 6A is a mass spectrum of 1,1,1-trichloroethane obtained in a conventional manner. The peak at mass 97 corresponds to $\text{CH}_3\text{CCl}_2^+$. In comparison, FIG. 6B is a mass spectrum of 1,1,1-trichloroethane obtained using the same experimental parameters as FIG. 6A, except that a low-frequency supplemental squarewave voltage (100 Hz, 42 volts) was applied for 20 milliseconds. It can be seen from FIG. 6B that the peak intensity at mass 97 has been reduced, and that ions of mass 61 (CH_2CCl^+) are abundant. As a result of non-resonant excitation, the mass 97 ions absorbed energy and some were dissociated to form the mass 61 ions.

FIGS. 6C and 6D show spectra of 1,1,1-trichloroethane obtained using the same parameters used to obtain the results of FIGS. 6A and 6B, except that the frequency of the supplemental squarewave was set at

300 and 600Hz, respectively. The similarity of the spectra of FIGS. 6B, 6C and 6D show that the dissociation is largely independent of the frequency of the supplemental field over a broad range. Finally, FIG. 6E shows a mass spectrum of 1,1,1-trichloroethane obtained using the method of the prior art, *i.e.*, rather than use a non-resonant low-frequency squarewave, a resonant sine wave of 139.6KHz (the z-axis resonant frequency of ion mass 97) was applied for 20ms at a level of 800 mv. It can be seen that the daughter ion yields of both methods were about the same.

FIGS. 7A-C show mass spectra of PFTBA under various conditions to demonstrate how the method of the present invention may be used to eliminate high mass ions from the ion trap. FIG. 7A shows a complete mass spectra including both the parent and fragment ions. FIG. 7B shows that all ions with mass above 131 were eliminated from the trap when the voltage of the supplemental squarewave was raised to 20v. FIG. 7C shows that raising the voltage to 33v causes all ions with mass greater than 100 to be eliminated from the trap.

The application to chemical ionization experiments of the ability to eliminate high mass ions from the ion trap by using a low frequency supplemental field is shown in FIGS. 8A-C. FIGS. 8A-C show the same CI experiments of FIGS. 4B, 4D and 4G, respectively. However, rather than using broadband resonance ejection to eliminate unwanted sample ions from the trap, a low frequency supplemental waveform was used. It can be seen that the results are substantially the same by either method. The FIG. 8A results were obtained using a supplemental field having a frequency of 600 Hz; the FIG. 8B results were obtained using a supplemental field having frequency of 300 Hz; and the FIG. 8C results were obtained using a supplemental field having frequency of 400 Hz. In each case the magnitude of the supplemental voltage was between 20 and 40 v.

While the present invention has been described in connection with the preferred embodiments thereof, such description is not intended to be limiting and other variations and equivalents will be readily apparent to those skilled in the art. Accordingly, the scope of the invention should be determined solely by reference to the following claims. For example, while the invention has been described, in part, in connection with the performance of chemical ionization experiments preceded by an electron impact ionization step, the method could also be performed using photoionization in lieu of electron impact ionization.

Claims

1. A method of adjusting the dynamic range of an ion trap mass spectrometer used in the chemical ionization multiple scan mode, comprising the steps of:

(a) applying a trapping field to said ion trap such that ions within a range of desired mass-

to-charge ratios will be stably trapped,

(b) introducing sample and reagent gas into the ion trap,

(c) ionizing said sample and reagent gas for an ionization period,

5

(d) removing sample ions formed during said ionization period from said trap,

(e) allowing sample molecules to react with said reagent ions for a chemical ionization period to form sample ions,

10

(f) scanning said trap to cause sample ions of sequential mass-to-charge ratios to leave the trap in order,

(g) detecting the sample ions as they leave the trap,

15

(h) identifying the sample ion that was present in the greatest concentration and determining the concentration of said sample ion,

(i) repeating steps (a) through (g) using said concentration information to adjust either the ionization period or the chemical ionization period or both.

20

25

30

35

40

45

50

55




FIG. 1




FIG. 2













FIG. 3










FIG.5






FIG. 6C




FIG. 6D








FIG. 6E







European Patent
Office

EUROPEAN SEARCH REPORT

Application Number
EP 97 10 4015

DOCUMENTS CONSIDERED TO BE RELEVANT			
Category	Citation of document with indication, where appropriate, of relevant passages	Relevant to claim	CLASSIFICATION OF THE APPLICATION (Int.Cl.6)
D,A	US 4 771 172 A (WEBER-GRABAU MICHAEL ET AL) 13 September 1988 * page 5, line 34 - page 6, line 5; claim 1; figure 3 *	1	H01J49/42
D,A	US 4 686 367 A (LOURIS JOHN N ET AL) 11 August 1987 * column 6, line 32 - line 68; claim 1; figure 4 *	1	
D,A	US 4 736 101 A (SYKA JOHN E P ET AL) 5 April 1988 * abstract *	1	
The present search report has been drawn up for all claims			TECHNICAL FIELDS SEARCHED (Int.Cl.6) H01J
Place of search THE HAGUE		Date of completion of the search 30 May 1997	Examiner Hulne, S
CATEGORY OF CITED DOCUMENTS X : particularly relevant if taken alone Y : particularly relevant if combined with another document of the same category A : technological background O : non-written disclosure P : intermediate document T : theory or principle underlying the invention E : earlier patent document, but published on, or after the filing date D : document cited in the application L : document cited for other reasons & : member of the same patent family, corresponding document			

EPO FORM 1503 03/92 (P/MCOU)