

(19)



(11)

EP 2 617 398 A1

(12)

EUROPEAN PATENT APPLICATION

(43) Date of publication:

24.07.2013 Bulletin 2013/30

(51) Int Cl.:

A61F 9/00 (2006.01)
A61F 2/16 (2006.01)

A61F 9/008 (2006.01)

(21) Application number: **13164193.8**

(22) Date of filing: **13.03.2008**

(84) Designated Contracting States:

**AT BE BG CH CY CZ DE DK EE ES FI FR GB GR
HR HU IE IS IT LI LT LU LV MC MT NL NO PL PT
RO SE SI SK TR**

(30) Priority: **13.03.2007 US 906944 P**

(62) Document number(s) of the earlier application(s) in
accordance with Art. 76 EPC:

08726825.6 / 2 129 346

(71) Applicant: **Optimedica Corporation**
Sunnyvale, CA 94089 (US)

(72) Inventor: **The designation of the inventor has not
yet been filed**

(74) Representative: **Gillard, Matthew Paul**
Withers & Rogers LLP
4 More London Riverside
London
SE1 2AU (GB)

Remarks:

This application was filed on 17-04-2013 as a
divisional application to the application mentioned
under INID code 62.

(54) **Intraocular lens for improved placement**

(57) A system and method for inserting an intraocular lens in a patient's eye includes a light source for generating a light beam, a scanner for deflecting the light beam to form an enclosed treatment pattern that includes a registration feature, and a delivery system for delivering the enclosed treatment pattern to target tissue in the pa-

tient's eye to form an enclosed incision therein having the registration feature. An intraocular lens is placed within the enclosed incision, wherein the intraocular lens has a registration feature that engages with the registration feature of the enclosed incision. Alternately, the scanner can make a separate registration incision for a post that is connected to the intraocular lens via a strut member.




Figure 14

EP 2 617 398 A1

Description

RELATED APPLICATIONS

[0001] This application claims the benefit of U.S. Provisional Application No. 60/906,944, filed March 13, 2007, and which is incorporated herein by reference.

FIELD OF THE INVENTION

[0002] The present invention relates to ophthalmic surgical procedures and systems.

BACKGROUND OF THE INVENTION

[0003] Cataract extraction is one of the most commonly performed surgical procedures in the world with estimated 2.5 million cases performed annually in the United States and 9.1 million cases worldwide in 2000. This was expected to increase to approximately 13.3 million estimated global cases in 2006. This market is composed of various segments including intraocular lenses for implantation, viscoelastic polymers to facilitate surgical maneuvers, disposable instrumentation including ultrasonic phacoemulsification tips, tubing, and various knives and forceps. Modern cataract surgery is typically performed using a technique termed phacoemulsification in which an ultrasonic tip with an associated water stream for cooling purposes is used to sculpt the relatively hard nucleus of the lens after performance of an opening in the anterior lens capsule termed anterior capsulotomy or more recently capsulorhexis. Following these steps as well as removal of residual softer lens cortex by aspiration methods without fragmentation, a synthetic foldable intraocular lens (IOL) is inserted into the eye through a small incision.

[0004] One of the earliest and most critical steps in the procedure is the performance of capsulorhexis. This step evolved from an earlier technique termed can-opener capsulotomy in which a sharp needle was used to perforate the anterior lens capsule in a circular fashion followed by the removal of a circular fragment of lens capsule typically in the range of 5 - 8 mm in diameter. This facilitated the next step of nuclear sculpting by phacoemulsification. Due to a variety of complications associated with the initial can-opener technique, attempts were made by leading experts in the field to develop a better technique for removal of the anterior lens capsule preceding the emulsification step. The concept of the capsulorhexis is to provide a smooth continuous circular opening through which not only the phacoemulsification of the nucleus can be performed safely and easily, but also for easy insertion of the intraocular lens. It provides both a clear central access for insertion, a permanent aperture for transmission of the image to the retina by the patient, and also a support of the IOL inside the remaining capsule that would limit the potential for dislocation.

[0005] Using the older technique of can-opener capsulotomy, or even with the continuous capsulorhexis, problems may develop related to inability of the surgeon to adequately visualize the capsule due to lack of red reflex, to grasp it with sufficient security, to tear a smooth circular opening of the appropriate size without radial rips and extensions or technical difficulties related to maintenance of the anterior chamber depth after initial opening, small size of the pupil, or the absence of a red reflex due to the lens opacity. Some of the problems with visualization have been minimized through the use of dyes such as methylene blue or indocyanine green. Additional complications arise in patients with weak zonules (typically older patients) and very young children that have very soft and elastic capsules, which are very difficult to mechanically rupture.

[0006] Many cataract patients are astigmatic. Astigmatism can occur when the cornea has a different curvature one direction than the other. IOLS are used for correcting astigmatism but require precise placement, orientation, and stability. Complete and long lasting correction using IOLs is difficult. Further, IOLs are not presently used to correct beyond 5D of astigmatism, even though many candidates have more severe aberrations. Correcting it further often involves making the corneal shape more spherical, or at least more radially symmetrical. There have been numerous approaches, including Corneal Relaxing Incisions (CRI), and Limbal Relaxing Incisions (LRI). All are done using manual, mechanical incisions. Presently, astigmatism cannot easily or predictably be fully corrected. About one third of those who have surgery to correct the irregularity find that their eyes regress to a considerable degree and only a small improvement is noted. Another third find that the astigmatism has been significantly reduced but not fully corrected. The remaining third have the most encouraging results with the most or all of the desired correction achieved.

[0007] What is needed are ophthalmic methods, techniques and apparatus to advance the standard of care of the astigmatic cataract patient.

SUMMARY OF THE INVENTION

[0008] The present invention provides methods and apparatus to precisely and accurately seat an IOL within the capsule of an eye of a patient by using a short pulse laser to create a capsular incision whose size and shape complement that of the IOL itself. This can be accomplished by adding asymmetrical features to the incision and lens, or portions thereof.

[0009] A device for treating a lesion on target tissue includes a light source for generating a light beam, a scanner for deflecting the light beam in the form of an enclosed treatment pattern that includes a registration feature, and a delivery system of one or more optical elements configured to deliver the enclosed treatment pattern to target tissue to form an enclosed incision there-

in having the registration feature.

[0010] An intraocular lens for insertion into a patient's eye includes a lens portion shaped for focusing light passing therethrough, and at least two haptics extending from the lens portion for supplying a spring force to the lens portion. The lens portion includes a peripheral edge that forms a registration feature.

[0011] An intraocular lens for insertion into a patient's eye includes a lens portion shaped for focusing light passing therethrough, at least two haptics extending from the lens portion for supplying a spring force to the lens portion, and a post separated from, and connected to the lens portion by, at least one strut member.

[0012] Other objects and features of the present invention will become apparent by a review of the specification, claims and appended figures.

BRIEF DESCRIPTION OF THE DRAWINGS

[0013] Fig. 1 is a schematic diagram of the optical beam scanning system.

[0014] Fig. 2 is an optical diagram showing an alternative beam combining scheme.

[0015] Fig. 3 is a schematic diagram of the optical beam scanning system with an alternative OCT configuration.

[0016] Fig. 4 is a schematic diagram of the optical beam scanning system with another alternative OCT combining scheme.

[0017] Fig. 5 is a top view diagram of a rotationally asymmetric capsulorhexis incision.

[0018] Fig. 6 is a top view diagram of a complementary rotationally asymmetric IOL.

[0019] Fig. 7 is a top view the IOL of Figure 6 positioned in the capsule of Figure 5.

[0020] Figs. 8 and 9 are side views of the rotationally asymmetric IOL of Figure 6.

[0021] Fig. 10 is a top view diagram of a rotationally asymmetric capsulorhexis incision.

[0022] Fig. 11 is a top view diagram of a complementary rotationally asymmetric IOL.

[0023] Fig. 12 is a top view of the IOL of Figure 11 positioned in the capsule of Figure 10.

[0024] Fig. 13 is a top view diagram of a rotationally asymmetric capsulorhexis incision.

[0025] Fig. 14 is a top view shows a diagram of a complementary rotationally asymmetric IOL.

[0026] Fig. 15 is a top view diagram of a rotationally asymmetric capsulorhexis incision.

[0027] Fig. 16 is a top view diagram of a complementary rotationally asymmetric IOL.

[0028] Fig. 17 is a perspective view of the rotationally asymmetric IOL of Figure 16.

[0029] Fig. 18 is a side view of the rotationally asymmetric IOL of Figure 16.

[0030] Fig. 19 is a top view diagram of a rotationally asymmetric IOL.

[0031] Fig. 20 is a top view diagram of a rotationally

asymmetric IOL.

DETAILED DESCRIPTION OF THE PREFERRED EMBODIMENTS

[0032] The techniques and systems disclosed herein provide many advantages over the current standard of care. Specifically, rapid and precise openings in the lens capsule are enabled using 3-dimensional patterned laser cutting to facilitate the placement and stability of intraocular lenses.

[0033] Another procedure enabled by the techniques described herein provides for the controlled formation of an incision in the anterior and/or posterior lens capsule. Conventional procedures require a complete circle or nearly complete circular cut. Openings formed using conventional, manual capsulorhexis techniques rely primarily on the mechanical shearing properties of lens capsule tissue and uncontrollable tears of the lens capsule to form openings. These conventional techniques are confined to the central lens portion or to areas accessible using mechanical cutting instruments and to varying limited degrees utilize precise anatomical measurements during the formation of the tears. In contrast, the controllable, patterned laser techniques described herein may be used to create an incision in virtually any position in the anterior and/or posterior capsule(s) and in virtually any shape. In "Bag-in-the-lens" surgery, matching incisions must be made in both the anterior and posterior capsules to fit the IOL in place. The present invention is uniquely suited to perform such matching incisions.

[0034] Furthermore, these capsular incisions may be tailored or keyed to accommodate an asymmetric IOL that requires it to be precisely positioned in both its location and rotational orientation. Moreover, the controllable, patterned laser techniques described herein also have available and/or utilize precise lens capsule size, measurement and other dimensional information that allows the incision or opening formation while minimizing impact on surrounding tissue.

[0035] The present invention can be implemented by a system that projects or scans an optical beam into a patient's eye 68, such as system 2 shown in Figure 1 which includes an ultrafast (UF) light source 4 (e.g. a femtosecond laser). Using this system, a beam may be scanned in a patient's eye in three dimensions: X, Y, Z. In this embodiment, the UF wavelength can vary between 1010nm to 1100nm and the pulse width can vary from 100fs to 10000fs. The pulse repetition frequency can also vary from 10kHz to 250kHz. Safety limits with regard to unintended damage to non-targeted tissue bound the upper limit with regard to repetition rate and pulse energy; while threshold energy, time to complete the procedure and stability bound the lower limit for pulse energy and repetition rate. The peak power of the focused spot in the eye 68 and specifically within the crystalline lens 69 and anterior capsule of the eye is sufficient to produce optical breakdown and initiate a plasma-mediated abla-

tion process. Near-infrared wavelengths are preferred because linear optical absorption and scattering in biological tissue is reduced across that spectral range. As an example, laser 4 may be a repetitively pulsed 1035 nm device that produces 500 fs pulses at a repetition rate of 100kHz and an individual pulse energy in the ten microjoule range.

[0036] The laser 4 is controlled by control electronics 300, via an input and output device 302, to create optical beam 6. Control electronics 300 may be a computer, microcontroller, etc. In this example, the entire system is controlled by the controller 300, and data moved through input/output device IO 302. A graphical user interface GUI 304 may be used to set system operating parameters, process user input (UI) 306 on the GUI 304, and display gathered information such as images of ocular structures.

[0037] The generated UF light beam 6 proceeds towards the patient eye 68 passing through half-wave plate, 8, and linear polarizer, 10. The polarization state of the beam can be adjusted so that the desired amount of light passes through half-wave plate 8 and linear polarizer 10, which together act as a variable attenuator for the UF beam 6. Additionally, the orientation of linear polarizer 10 determines the incident polarization state incident upon beamcombiner 34, thereby optimizing beamcombiner throughput.

[0038] The UF beam proceeds through a shutter 12, aperture 14, and a pickoff device 16. The system controlled shutter 12 ensures on/off control of the laser for procedural and safety reasons. The aperture sets an outer useful diameter for the laser beam and the pickoff monitors the output of the useful beam. The pickoff device 16 includes of a partially reflecting mirror 20 and a detector 18. Pulse energy, average power, or a combination may be measured using detector 18. The information can be used for feedback to the half-wave plate 8 for attenuation and to verify whether the shutter 12 is open or closed. In addition, the shutter 12 may have position sensors to provide a redundant state detection.

[0039] The beam passes through a beam conditioning stage 22, in which beam parameters such as beam diameter, divergence, circularity, and astigmatism can be modified. In this illustrative example, the beam conditioning stage 22 includes a 2 element beam expanding telescope comprised of spherical optics 24 and 26 in order to achieve the intended beam size and collimation. Although not illustrated here, an anamorphic or other optical system can be used to achieve the desired beam parameters. The factors used to determine these beam parameters include the output beam parameters of the laser, the overall magnification of the system, and the desired numerical aperture (NA) at the treatment location. In addition, the optical system 22 can be used to image aperture 14 to a desired location (e.g. the center location between the 2-axis scanning device 50 described below). In this way, the amount of light that makes it through the aperture 14 is assured to make it through

the scanning system. Pickoff device 16 is then a reliable measure of the usable light.

[0040] After exiting conditioning stage 22, beam 6 reflects off of fold mirrors 28, 30, & 32. These mirrors can be adjustable for alignment purposes. The beam 6 is then incident upon beam combiner 34. Beamcombiner 34 reflects the UF beam 6 (and transmits both the OCT 114 and aim 202 beams described below). For efficient beamcombiner operation, the angle of incidence is preferably kept below 45 degrees and the polarization where possible of the beams is fixed. For the UF beam 6, the orientation of linear polarizer 10 provides fixed polarization.

[0041] Following the beam combiner 34, the beam 6 continues onto the z-adjust or Z scan device 40. In this illustrative example the z-adjust includes a Galilean telescope with two lens groups 42 and 44 (each lens group includes one or more lenses). Lens group 42 moves along the z-axis about the collimation position of the telescope. In this way, the focus position of the spot in the patient's eye 68 moves along the z-axis as indicated. In general there is a fixed linear relationship between the motion of lens 42 and the motion of the focus. In this case, the z-adjust telescope has an approximate 2x beam expansion ratio and a 1:1 relationship of the movement of lens 42 to the movement of the focus. Alternatively, lens group 44 could be moved along the z-axis to actuate the z-adjust, and scan. The z-adjust is the z-scan device for treatment in the eye 68. It can be controlled automatically and dynamically by the system and selected to be independent or to interplay with the X-Y scan device described next. Mirrors 36 and 38 can be used for aligning the optical axis with the axis of z-adjust device 40.

[0042] After passing through the z-adjust device 40, the beam 6 is directed to the x-y scan device by mirrors 46 & 48. Mirrors 46 & 48 can be adjustable for alignment purposes. X-Y scanning is achieved by the scanning device 50 preferably using two mirrors 52 & 54 under the control of control electronics 300, which rotate in orthogonal directions using motors, galvanometers, or any other well known optic moving device. Mirrors 52 & 54 are located near the telecentric position of the objective lens 58 and contact lens 66 combination described below. Tilting these mirrors 52/54 causes them to deflect beam 6, causing lateral displacements in the plane of UF focus located in the patient's eye 68. Objective lens 58 may be a complex multi-element lens element, as shown, and represented by lenses 60, 62, and 64. The complexity of the lens 58 will be dictated by the scan field size, the focused spot size, the available working distance on both the proximal and distal sides of objective 58, as well as the amount of aberration control. An f-theta lens 58 of focal length 60mm generating a spot size of 10 μ m, over a field of 10mm, with an input beam size of 15mm diameter is an example. Alternatively, X-Y scanning by scanner 50 may be achieved by using one or more moveable optical elements (e.g. lenses, gratings) which also may be controlled by control electronics 300, via input and

output device 302.

[0043] The aiming and treatment scan patterns can be automatically generated by the scanner 50 under the control of controller 300. Such patterns may be comprised of a single spot of light, multiple spots of light, a continuous pattern of light, multiple continuous patterns of light, and/or any combination of these. In addition, the aiming pattern (using aim beam 202 described below) need not be identical to the treatment pattern (using light beam 6), but preferably at least defines its boundaries in order to assure that the treatment light is delivered only within the desired target area for patient safety. This may be done, for example, by having the aiming pattern provide an outline of the intended treatment pattern. This way the spatial extent of the treatment pattern may be made known to the user, if not the exact locations of the individual spots themselves, and the scanning thus optimized for speed, efficiency and accuracy. The aiming pattern may also be made to be perceived as blinking in order to further enhance its visibility to the user.

[0044] An optional contact lens 66, which can be any suitable ophthalmic lens, can be used to help further focus the optical beam 6 into the patient's eye 68 while helping to stabilize eye position. The positioning and character of optical beam 6 and/or the scan pattern the beam 6 forms on the eye 68 may be further controlled by use of an input device such as a joystick, or any other appropriate user input device (e.g. GUI 304) to position the patient and/or the optical system.

[0045] The UF laser 4 and controller 300 can be set to target the surfaces of the targeted structures in the eye 68 and ensure that the beam 6 will be focused where appropriate and not unintentionally damage non-targeted tissue. Imaging modalities and techniques described herein, such as for example, Optical Coherence Tomography (OCT), Purkinje imaging, Scheimpflug imaging, or ultrasound may be used to determine the location and measure the thickness of the lens and lens capsule to provide greater precision to the laser focusing methods, including 2D and 3D patterning. Laser focusing may also be accomplished using one or more methods including direct observation of an aiming beam, Optical Coherence Tomography (OCT), Purkinje imaging, Scheimpflug imaging, ultrasound, or other known ophthalmic or medical imaging modalities and/or combinations thereof. In the embodiment of Figure 1, an OCT device 100 is described, although other modalities are within the scope of the present invention. An OCT scan of the eye will provide information about the axial location of the anterior and posterior lens capsule, the boundaries of the cataract nucleus, as well as the depth of the anterior chamber. This information is then be loaded into the control electronics 300, and used to program and control the subsequent laser-assisted surgical procedure. The information may also be used to determine a wide variety of parameters related to the procedure such as, for example, the upper and lower axial limits of the focal planes used for cutting the lens capsule and segmentation of the lens

cortex and nucleus, and the thickness of the lens capsule among others.

[0046] The OCT device 100 in Figure 1 includes a broadband or a swept light source 102 that is split by a fiber coupler 104 into a reference arm 106 and a sample arm 110. The reference arm 106 includes a module 108 containing a reference reflection along with suitable dispersion and path length compensation. The sample arm 110 of the OCT device 100 has an output connector 112 that serves as an interface to the rest of the UF laser system. The return signals from both the reference and sample arms 106, 110 are then directed by coupler 104 to a detection device 128, which employs either time domain, frequency or single point detection techniques. In Figure 1, a frequency domain technique is used with an OCT wavelength of 920nm and bandwidth of 100nm.

[0047] Exiting connector 112, the OCT beam 114 is collimated using lens 116. The size of the collimated beam 114 is determined by the focal length of lens 116. The size of the beam 114 is dictated by the desired NA at the focus in the eye and the magnification of the beam train leading to the eye 68. Generally, OCT beam 114 does not require as high an NA as the UF beam 6 in the focal plane and therefore the OCT beam 114 is smaller in diameter than the UF beam 6 at the beamcombiner 34 location. Following collimating lens 116 is aperture 118 which further modifies the resultant NA of the OCT beam 114 at the eye. The diameter of aperture 118 is chosen to optimize OCT light incident on the target tissue and the strength of the return signal. Polarization control element 120, which may be active or dynamic, is used to compensate for polarization state changes which may be induced by individual differences in corneal birefringence, for example. Mirrors 122 & 124 are then used to direct the OCT beam 114 towards beamcombiners 126 & 34. Mirrors 122 & 124 may be adjustable for alignment purposes and in particular for overlaying of OCT beam 114 to UF beam 6 subsequent to beamcombiner 34. Similarly, beamcombiner 126 is used to combine the OCT beam 114 with the aim beam 202 described below.

[0048] Once combined with the UF beam 6 subsequent to beamcombiner 34, OCT beam 114 follows the same path as UF beam 6 through the rest of the system. In this way, OCT beam 114 is indicative of the location of UF beam 6. OCT beam 114 passes through the z-scan 40 and x-y scan 50 devices then the objective lens 58, contact lens 66 and on into the eye 68. Reflections and scatter off of structures within the eye provide return beams that retrace back through the optical system, into connector 112, through coupler 104, and to OCT detector 128. These return back reflections provide the OCT signals that are in turn interpreted by the system as to the location in X, Y Z of UF beam 6 focal location.

[0049] OCT device 100 works on the principle of measuring differences in optical path length between its reference and sample arms. Therefore, passing the OCT through z-adjust 40 does not extend the z-range of OCT system 100 because the optical path length does not

change as a function of movement of 42. OCT system 100 has an inherent z-range that is related to the detection scheme, and in the case of frequency domain detection it is specifically related to the spectrometer and the location of the reference arm 106. In the case of OCT system 100 used in Figure 1, the z-range is approximately 1-2mm in an aqueous environment. Extending this range to at least 4mm involves the adjustment of the path length of the reference arm within OCT system 100. Passing the OCT beam 114 in the sample arm through the z-scan of z-adjust 40 allows for optimization of the OCT signal strength. This is accomplished by focusing the OCT beam 114 onto the targeted structure while accommodating the extended optical path length by commensurately increasing the path within the reference arm 106 of OCT system 100.

[0050] Because of the fundamental differences in the OCT measurement with respect to the UF focus device due to influences such as immersion index, refraction, and aberration, both chromatic and monochromatic, care must be taken in analyzing the OCT signal with respect to the UF beam focal location. A calibration or registration procedure as a function of X, Y Z should be conducted in order to match the OCT signal information to the UF focus location and also to the relate to absolute dimensional quantities.

[0051] Observation of an aim beam may also be used to assist the user to directing the UF laser focus. Additionally, an aim beam visible to the unaided eye in lieu of the infrared OCT and UF beams can be helpful with alignment provided the aim beam accurately represents the infrared beam parameters. An aim subsystem 200 is employed in the configuration shown in Figure 1. The aim beam 202 is generated by a an aim beam light source 201, such as a helium-neon laser operating at a wavelength of 633nm. Alternatively a laser diode in the 630-650nm range could be used. The advantage of using the helium neon 633nm beam is its long coherence length, which would enable the use of the aim path as a laser unequal path interferometer (LUPI) to measure the optical quality of the beam train, for example.

[0052] Once the aim beam light source generates aim beam 202, the aim beam 202 is collimated using lens 204. The size of the collimated beam is determined by the focal length of lens 204. The size of the aim beam 202 is dictated by the desired NA at the focus in the eye and the magnification of the beam train leading to the eye 68. Generally, aim beam 202 should have close to the same NA as UF beam 6 in the focal plane and therefore aim beam 202 is of similar diameter to the UF beam at the beamcombiner 34 location. Because the aim beam is meant to stand-in for the UF beam 6 during system alignment to the target tissue of the eye, much of the aim path mimics the UF path as described previously. The aim beam 202 proceeds through a half-wave plate 206 and linear polarizer 208. The polarization state of the aim beam 202 can be adjusted so that the desired amount of light passes through polarizer 208. Elements 206 &

208 therefore act as a variable attenuator for the aim beam 202. Additionally, the orientation of polarizer 208 determines the incident polarization state incident upon beamcombiners 126 and 34, thereby fixing the polarization state and allowing for optimization of the beamcombiners' throughput. Of course, if a semiconductor laser is used as aim beam light source 200, the drive current can be varied to adjust the optical power.

[0053] The aim beam 202 proceeds through a shutter 210 and aperture 212. The system controlled shutter 210 provides on/off control of the aim beam 202. The aperture 212 sets an outer useful diameter for the aim beam 202 and can be adjusted appropriately. A calibration procedure measuring the output of the aim beam 202 at the eye can be used to set the attenuation of aim beam 202 via control of polarizer 206.

[0054] The aim beam 202 next passes through a beam conditioning device 214. Beam parameters such as beam diameter, divergence, circularity, and astigmatism can be modified using one or more well known beaming conditioning optical elements. In the case of an aim beam 202 emerging from an optical fiber, the beam conditioning device 214 can simply include a beam expanding telescope with two optical elements 216 and 218 in order to achieve the intended beam size and collimation. The final factors used to determine the aim beam parameters such as degree of collimation are dictated by what is necessary to match the UF beam 6 and aim beam 202 at the location of the eye 68. Chromatic differences can be taken into account by appropriate adjustments of beam conditioning device 214. In addition, the optical system 214 is used to image aperture 212 to a desired location such as a conjugate location of aperture 14.

[0055] The aim beam 202 next reflects off of fold mirrors 222 & 220, which are preferably adjustable for alignment registration to UF beam 6 subsequent to beam combiner 34. The aim beam 202 is then incident upon beam combiner 126 where the aim beam 202 is combined with OCT beam 114. Beamcombiner 126 reflects the aim beam 202 and transmits the OCT beam 114, which allows for efficient operation of the beamcombining functions at both wavelength ranges. Alternatively, the transmit and reflect functions of beamcombiner 126 can be reversed and the configuration inverted. Subsequent to beamcombiner 126, aim beam 202 along with OCT beam 114 is combined with UF beam 6 by beamcombiner 34.

[0056] A device for imaging the target tissue on or within the eye 68 is shown schematically in Figure 1 as imaging system 71. Imaging system includes a camera 74 and an illumination light source 86 for creating an image of the target tissue. The imaging system 71 gathers images which may be used by the system controller 300 for providing pattern centering about or within a predefined structure. The illumination light source 86 for the viewing is generally broadband and incoherent. For example, light source 86 can include multiple LEDs as shown. The wavelength of the viewing light source 86 is preferably in the range of 700nm to 750nm, but can be

anything which is accommodated by the beamcombiner 56, which combines the viewing light with the beam path for UF beam 6 and aim beam 202 (beamcombiner 56 reflects the viewing wavelengths while transmitting the OCT and UF wavelengths). The beamcombiner 56 may partially transmit the aim wavelength so that the aim beam 202 can be visible to the viewing camera 74. Optional polarization element 84 in front of light source 86 can be a linear polarizer, a quarter wave plate, a half-wave plate or any combination, and is used to optimize signal. A false color image as generated by the near infrared wavelength is acceptable.

[0057] The illumination light from light source 86 is directed down towards the eye using the same objective lens 58 and contact lens 66 as the UF and aim beam 6, 202. The light reflected and scattered off of various structures in the eye 68 are collected by the same lenses 58 & 66 and directed back towards beamcombiner 56. There, the return light is directed back into the viewing path via beam combiner and mirror 82, and on to camera 74. Camera 74 can be, for example but not limited to, any silicon based detector array of the appropriately sized format. Video lens 76 forms an image onto the camera's detector array while optical elements 80 & 78 provide polarization control and wavelength filtering respectively. Aperture or iris 81 provides control of imaging NA and therefore depth of focus and depth of field. A small aperture provides the advantage of large depth of field which aids in the patient docking procedure. Alternatively, the illumination and camera paths can be switched. Furthermore, aim light source 200 can be made to emit in the infrared which would not directly visible, but could be captured and displayed using imaging system 71.

[0058] Coarse adjust registration is usually needed so that when the contact lens 66 comes into contact with the cornea, the targeted structures are in the capture range of the X, Y scan of the system. Therefore a docking procedure is preferred, which preferably takes in account patient motion as the system approaches the contact condition (i.e. contact between the patient's eye 68 and the contact lens 66. The viewing system 71 is configured so that the depth of focus is large enough such that the patient's eye 68 and other salient features may be seen before the contact lens 66 makes contact with eye 68.

[0059] Preferably, a motion control system 70 is integrated into the overall control system 2, and may move the patient, the system 2 or elements thereof, or both, to achieve accurate and reliable contact between contact lens 66 and eye 68. Furthermore, a vacuum suction subsystem and flange may be incorporated into system 2, and used to stabilize eye 68. The alignment of eye 68 to system 2 via contact lens 66 may be accomplished while monitoring the output of imaging system 71, and performed manually or automatically by analyzing the images produced by imaging system 71 electronically by means of control electronics 300 via IO 302. Force and/or pressure sensor feedback may also be used to discern contact, as well as to initiate the vacuum subsystem.

[0060] An alternative beamcombining configuration is shown in the alternate embodiment of Figure 2. For example, the passive beamcombiner 34 in Figure 1 can be replaced with an active combiner 140 in Figure 2. The active beamcombiner 34 can be a moving or dynamically controlled element such as a galvanometric scanning mirror, as shown. Active combiner 140 changes its angular orientation in order to direct either the UF beam 6 or the combined aim and OCT beams 202, 114 towards the scanner 50 and eventually eye 68 one at a time. The advantage of the active combining technique is that it avoids the difficulty of combining beams with similar wavelength ranges or polarization states using a passive beam combiner. This ability is traded off against the ability to have simultaneous beams in time and potentially less accuracy and precision due to positional tolerances of active beam combiner 140.

[0061] Another alternate embodiment is shown in Figure 3 which is similar to that of Figure 1 but utilizes an alternate approach to OCT 100. In Figure 3, OCT 101 is the same as OCT 100 in Figure 1, except that the reference arm 106 has been replaced by reference arm 132. This free-space OCT reference arm 132 is realized by including beamsplitter 130 after lens 116. The reference beam 132 then proceeds through polarization controlling element 134 and then onto the reference return module 136. The reference return module 136 contains the appropriate dispersion and path length adjusting and compensating elements and generates an appropriate reference signal for interference with the sample signal. The sample arm of OCT 101 now originates subsequent to beamsplitter 130. The potential advantages of this free space configuration include separate polarization control and maintenance of the reference and sample arms. The fiber based beam splitter 104 of OCT 101 can also be replaced by a fiber based circulator. Alternately, both OCT detector 128 and beamsplitter 130 might be moved together as opposed to reference arm 136.

[0062] Figure 4 shows another alternative embodiment for combining OCT beam 114 and UF beam 6. In Figure 4, OCT 156 (which can include either of the configurations of OCT 100 or 101) is configured such that its OCT beam 154 is coupled to UF beam 6 after the z-scan 40 using beamcombiner 152. In this way, OCT beam 154 avoids using the z-adjust. This allows the OCT 156 to possibly be folded into the beam more easily and shortening the path length for more stable operation. This OCT configuration is at the expense of an optimized signal return strength as discussed with respect to Figure 1. There are many possibilities for the configuration of the OCT interferometer, including time and frequency domain approaches, single and dual beam methods, swept source, etc, as described in U.S. Pat. Nos. 5,748,898; 5,748,352; 5,459,570; 6,111,645; and 6,053,613 (which are incorporated herein by reference.)

[0063] Figures 5 through 9 illustrate different aspects of an embodiment of the present invention, which can be implemented using the scanning system 2 described

above. As shown in Figure 5, a capsulorhexis incision 400 (which may be created using system 2) is tailored for astigmatism-correcting intraocular lenses (IOLs). Such astigmatism-correcting IOLs need to be placed not only at the correct location within the capsule 402 of the eye 68, but also oriented at the correct rotational/clocking angle. Thus, they have inherent rotational asymmetries, unlike spherical IOLs. The incision 400 shown in this example is elliptical, however, other shapes are also useful. Incision 400 may be made continuously, or piecewise to largely maintain the structural integrity of the lens-capsule apparatus of the patient's eye 68. Such incomplete incisions 400 may be thought of as perforated incisions, and may be made to be removed gently in order to minimize their potential to inadvertently extend the capsulorhexis. Either way, incision 400 is an enclosed incision, which for the purposes of this disclosure means that it starts and ends at the same location and encircles a certain amount of tissue therein. The simplest example of an enclosed incision is a circular incision, where a round piece of tissue is encircled by the incision. It follows therefore that an enclosed treatment pattern (i.e. generated by system 2 for forming an enclosed incision) is one that also starts and ends at the same location and defines a space encircled thereby.

[0064] One key feature of the enclosed incision 400 is that it includes a registration feature to orient the IOL that will be placed inside it. For the illustrated elliptical incision 400, its elliptical shape is its registration feature, which allows for the accurate placement of an IOL by virtue of its inherent rotational asymmetry, unlike the desired circular outcome of a manual CCC. The elliptical major axis 404 and minor axis 406 of incision 400 are shown. Major axis 404 and minor axis 406 are not equal. Incision 400 may be made at any rotational angle relative to the eye 68 of a patient, although it is shown in this example to be in the plane of the iris with its major axis 404 lying along the horizontal. Incision 400 is intended to mate with one or more complementary registration features on an IOL. The ranging subsystem of system 2 (e.g. the OCT 100 subsystem) may be used to precisely define the surface of the capsule 402 to be incised. This may serve to isolate the laser pulses nominally to the vicinity of the targeted capsule 402 itself, thus minimizing the energy required and the treatment time and commensurately increasing patient safety and overall efficiency.

[0065] As shown in Figure 6, an IOL 408 includes an optic portion 410 used to focus light and a haptic 416 used to position the IOL 408. Optic 410 is a rotationally asymmetric lens (about its optical axis) that include an elliptically shaped peripheral sidewall or edge 412, the complementary registration feature that mates with elliptically shaped incision 400. In this example, the elliptically shaped edge 412 includes a major axis 418 and minor axis 420. Major axis 418 and minor axis 420 are not equal. Intraocular lens IOL 408 further contains surface 414 that serves to hold haptics element 416 and provide a resting place for capsule 402 to secure optic 410 of intraocular

lens 408 in the proper orientation and position within the capsule 402 of a patient's eye 68. Surface 414 is shown as elliptical, but need not be. Haptics 416 provide stability and may serve to seat edge 412 of intraocular lens 408 in incision 400 by applying retaining force towards the anterior portion of capsule 402. Haptics 416 may be deployed in any orientation. The orientation of the cylindrical correction of optic 410 of intraocular lens 408 may be made to coincide with either its major axis 418 or its minor axis 420. In this way, intraocular lenses IOL 408 and optic 410 may be manufactured in a standardized manner and the rotational orientation of incision 400 and the spherical and cylindrical optical powers of optic 410 may be made to vary to suit the individual optical prescription of the eye 68 of a patient.

[0066] Figure 7 shows the proper immediate disposition of intraocular lens 408 once installed into capsule 402 with mating registration features edge 412 and incision 400 engaged, and resting upon surface 414. Major axis 404 and major axis 418 are not of equal length. Minor axis 406 and minor axis 420 are not the same length, either. This is done to accommodate the fact the capsule 402 may contract somewhat subsequent to capsulorhexis incision. The difference between the lengths of these axes is intended to allow the capsule 402 to contract and still better seat intraocular lens 408 into capsule 402 via incision 400. These differences should be limited to allow for reasonable contraction, but not so much as to allow for significant rotation of intraocular lens 408. Typical values for these length differences may range from 100 μ m to 500 μ m, for example.

[0067] Figure 8 shows a side view on the same intraocular lens 408 depicted in Figures 6 and 7. In this schematic representation, edge 412 is shown on the same side of optic 410 as surface 424 of intraocular lens 408. The surface 422 on intraocular lens 408 serves to maintain the integrity of fit between edge 412 and incision 400. Edge 412 is seen as the projection of surface 422 in the alternate view depicted in Figures 6 and 7. Optical axis 411 of optic 410 is shown. Haptics 416 lie along the line of sight in this view.

[0068] Figure 9 is a side view of the lens configuration of Figure 8, but rotated 90 degrees to show that displaying surface 426 is not curved in both directions (i.e. shaped as a cylindrical lens). This cylindrical or toric optical system of optic 410 provides cylindrical correction for the astigmatism of a patient. Haptics 416 lie perpendicular to the line of sight in this view.

[0069] Figure 10 shows an alternate embodiment similar to the asymmetry of the example of Figure 6, except that incision 400 includes a registration feature 428 formed as a notch extending from the otherwise round incision 400. Registration feature 428 serves to provide a means to locate a matching registration feature (i.e. a protrusion) on intraocular lens 410. The complementary registration feature 430 of IOL 408 including optic 410 is illustrated in Figure 11. The shape of registration feature 428 is shown as half round for illustrative purposes only.

Alternately, a teardrop shape, such as that shown in Figure 15, for edge 412 and incision 400 is less likely to contain sharp edges and thus be less prone to inadvertent extension of the capsulorhexis. Many similar complementary shapes are possible and within the scope of the present invention. A benefit of the short pulse laser system described in Figure 1 is that it may provide via a plasma-mediated ablation process smooth incisions 400 that are unlikely to extend.

[0070] In Figure 11, registration feature 430 is intended to mate with registration feature 428 of incision 400. This serves to correctly locate optic 410 and maintain its rotational integrity. Here again edge 412 and surface 414 provide features to assure mechanical stability and proper orientation with respect to the capsule 402 of a patient's eye 68. Similar to the Figure 6 description of the asymmetric major axes 404 & 418, registration feature 430 may be placed at an arbitrary rotational orientation to suit an individual prescription. Haptics 416 may be deployed in any orientation, as before.

[0071] Figure 12 shows the proper immediate disposition of intraocular lens 408 once installed into capsule 402 via incision 400 with mating feature edge 412 engaged, similar to that shown in Figure 7.

[0072] Figure 13 shows an alternate embodiment similar to that of Figures 6 & 10 with the addition of a registration incision 432 (formed by a registration pattern of treatment light generated by system 2) that is separate and distinct from capsulorhexis incision 400. As before, registration incision 432 serves to provide a means to locate a matching registration feature on intraocular lens 408.

[0073] Figure 14 shows an alternate embodiment similar to that of Figure 11, with post 434 that sits in the registration incision 432 and atop strut 436 away from optic 410 on intraocular lens 408. Post 434 and strut 436 are shown as being tilted away from the normal between haptics 416, but need not be. Many such similar complementary configurations are possible and within the scope of the present invention.

[0074] IOL 408 can also mate with the capsulorhexis incision by way of a circumferential flange. The shape of the capsulorhexis incision 400 may be made to orient the IOL 408 to achieve cylindrical corrections, as shown schematically in Figure 15. The asymmetric incision 400 of Figure 15 is similar to that of Figures 5, 10 and 13 with the addition that it is intended to mate with a flange on intraocular lens 408 rather than an edge 412 and a surface 414.

[0075] Figure 16 shows intraocular lens 408 utilizing a flange 438 to mate with incision 400. As shown, intraocular lens 408 is comprised of optic 410 and flange 438. This flange 438 may be circumferential, as shown, but need not be. It may simply lie atop optic 410 and serve the same purpose of mating and retaining intraocular lens 408 within capsule 402. Flange 438 contains groove 440 to seat the capsule 402 in incision 400. Rotationally asymmetric groove 440 serves to accurately position and

retain intraocular lens 408 within incision 400 at the correct rotational orientation for the individual astigmatic prescription. This optical correction is achieved using optic 410. Alternately, groove 440 may be created between flange 438 and optic 410 (rather than within flange 438, as shown) when flange 438 lies atop intraocular lens 408. Such an intraocular lens 408 may be used in "bag-in-the-lens" surgeries.

[0076] Figures 17 and 18 show the same configuration as that of Figure 16, but from different viewing perspectives to better illustrate groove 440. Groove 440 may be made to engage incision 400 continuously, as shown, or discontinuously by providing notches cut into flange 438. Such notches may serve to more easily initiate the engagement of flange 438 with capsule 402 via incision 400. Alternately, flange 438 could be made such that the depth between its edge and groove 440 varies along its circumference. This way, a region of shallow depth could be used as a starting point for more easily engaging intraocular lens 408 with capsule 402 via incision 400.

[0077] Figure 19 shows an alternate embodiment that is similar to that of Figure 16, but where optic 410 may be made to rotate within flange 438. To align the rotation of optic 410, angular alignment marks 444 are displayed on flange 438 and a complementary alignment mark 446 is displayed on optic 410. In this manner, intraocular lens 408 and optic 410 may be manufactured in a standardized manner and one may rotate optic 410 relative to its surrounding flange 438 to provide astigmatic correction to suit the individual prescription. Alignment marks 444 are shown at 22.5° intervals, but may be otherwise. Alignment marks 444 and 446 may be etched into the materials of their host elements, or alternately imprinted upon them.

[0078] Figure 20 shows one further alternate embodiment that is similar to that of Figures 14 and 19, where optic 410 may be made to rotate within ring 448. In this illustrative example, post 434 and strut 436 are integral to optic 410, and contain alignment mark 446. Ring 448 contains haptics 416 and surface 414 as before but now also alignment marks 444. Alignment mark 446 on strut 436 of optic 410 facilitates the rotational orientation of astigmatic correcting optic 410. In this manner, the ultimate orientation of intraocular lens 408 within capsule 402 of the eye of a patient via incision 400 that may be made in any orientation. Many such similar complementary configurations are possible and within the scope of the present invention.

[0079] It is to be understood that the present invention is not limited to the embodiment(s) described above and illustrated herein, but encompasses any and all variations falling within the scope of the appended claims. For example, references to the present invention herein are not intended to limit the scope of any claim or claim term, but instead merely make reference to one or more features that may be covered by one or more of the claims. All the optical elements downstream of scanner 50 shown in Figs. 1, 3 and 4 form a delivery system of optical ele-

ments for delivering the beam 6,114 and 202 to the target tissue. Conceivably, depending on the desired features of the system, some or even most of the depicted optical elements could be omitted in a delivery system that still reliably delivers the scanned beams to the target tissue. Protrusion registration features could be replaced with indentations (i.e. notches), and vice versa, Before going on to present the claims, we will first set out a group of clauses drawing attention to some salient embodiments of the invention, as follows:

1. A device for treating a lesion on target tissue, comprising:

a light source for generating a light beam;
a scanner for deflecting the light beam in the form of an enclosed treatment pattern that includes a registration feature; and
a delivery system of one or more optical elements configured to deliver the enclosed treatment pattern to target tissue to form an enclosed incision therein having the registration feature.

2. The device of clause 1 wherein the registration feature is an asymmetric shape of the enclosed treatment pattern.

3. The device of clause 1 wherein the registration feature is a protrusion in an outer boundary of the enclosed treatment pattern.

4. The device of clause 1 wherein the registration feature is a notch in an outer boundary of the enclosed treatment pattern.

5. An intraocular lens for insertion into a patient's eye, comprising:

a lens portion shaped for focusing light passing therethrough; and
at least two haptics extending from the lens portion for supplying a spring force to the lens portion;
wherein the lens portion includes a peripheral edge that forms a registration feature.

6. The intraocular lens of clause 5 wherein the registration feature is an asymmetric shape of the peripheral edge.

7. The intraocular lens of clause 5 wherein the registration feature is a protrusion formed in the peripheral edge.

8. The intraocular lens of clause 5 wherein the registration feature is a notch

9. The intraocular lens of clause 5 wherein the pe-

ripheral edge includes a groove for engaging with tissue.

10. The intraocular lens of clause 5, further comprising:

a flange surrounding the lens portion.

11. The intraocular lens of clause 10, wherein the lens portion is configured to rotate within the flange.

12. The intraocular lens of clause 11, wherein at least one of the lens portion and flange includes angular alignment marks formed thereon.

13. An intraocular lens for insertion into a patient's eye, comprising:

a lens portion shaped for focusing light passing therethrough;
at least two haptics extending from the lens portion for supplying a spring force to the lens portion; and
a post separated from, and connected to the lens portion by, at least one strut member.

Claims

1. An intraocular lens for replacing a natural lens in a lens capsule of a patient's eye, the intraocular lens comprising:

a lens portion shaped for focusing light passing therethrough;
at least two haptics extending from the lens portion for supplying a spring force to the lens portion; and
a post separated from and connected to the lens portion by a strut member, the post being shaped to be received in a registration incision in the lens capsule so as to orient the intraocular lens.

2. The intraocular lens of claim 1, further comprising at least two haptics extending from the lens portion for supplying a spring force to the lens portion.

3. The intraocular lens of claim 1 or claim 2, wherein the post is disposed atop the strut member.

4. The intraocular lens of any of the preceding claims, wherein the lens portion comprises an astigmatic lens.

5. The intraocular lens of any of the preceding claims,

wherein:

the intraocular lens comprises a flange surrounding the lens portion; and
the lens portion can be rotated relative to the flange to facilitate rotational orientation of the astigmatic lens. 5

6. The intraocular lens of any of the preceding claims, wherein the registration feature is separate and distinct from a capsulorhexis incision. 10
7. The intraocular lens of any of the preceding claims, wherein the lens portion has an edge configured to be seated in a capsulorhexis incision and the intraocular lens further comprises an outer portion surrounding the edge and providing a surface that provides a resting place for the lens capsule. 15
8. The intraocular lens of claim 2, wherein the lens portion has an edge configured to be seated in a capsulorhexis incision and the intraocular lens further comprises an outer portion surrounding the edge and providing a surface that provides a resting place for the lens capsule. 20
25
9. The intraocular lens of claim 8, wherein the at least two haptics are configured to seat the edge in the capsulorhexis incision by apply a retaining force to the lens portion towards the anterior portion of the lens capsule. 30

35

40

45

50

55




Figure 1




Figure 2




Figure 3




Figure 4




Figure 5.




Figure 6




Figure 7




Figure 8




Figure 9




Figure 10




Figure 11




Figure 12




Figure 13




Figure 14




Figure 15




Figure 16




Figure 17




Figure 18




Figure 19




Figure 20



EUROPEAN SEARCH REPORT

Application Number
EP 13 16 4193

DOCUMENTS CONSIDERED TO BE RELEVANT			
Category	Citation of document with indication, where appropriate, of relevant passages	Relevant to claim	CLASSIFICATION OF THE APPLICATION (IPC)
X	WO 94/07435 A1 (NESTLE SA [CH]; PATEL ANILBHAI S [US]) 14 April 1994 (1994-04-14)	1-4	INV. A61F9/00 A61F9/008 A61F2/16
Y	* page 13, line 3 - line 12 * * page 10, line 27 - line 33 * * figure 28 *	5-9	
X	----- US 2006/041308 A1 (NICHAMIN LOUIS D [US]) 23 February 2006 (2006-02-23) * abstract * * paragraph [0045] * * figure 13 *	1-3	
Y	----- US 2005/187623 A1 (TASSIGNON MARIE-JOSE B [BE]) 25 August 2005 (2005-08-25) * abstract * * paragraphs [0018], [0034], [0041] * * claim 6 *	6-9	
Y	----- US 2004/148022 A1 (EGGLESTON HARRY C [US]) 29 July 2004 (2004-07-29) * abstract * * figures 9b,12 * * paragraphs [0068], [0070], [0077], [0082] *	5	
The present search report has been drawn up for all claims			TECHNICAL FIELDS SEARCHED (IPC) A61F
Place of search Munich		Date of completion of the search 17 June 2013	Examiner Jansen, Birte
CATEGORY OF CITED DOCUMENTS X : particularly relevant if taken alone Y : particularly relevant if combined with another document of the same category A : technological background O : non-written disclosure P : intermediate document T : theory or principle underlying the invention E : earlier patent document, but published on, or after the filing date D : document cited in the application L : document cited for other reasons & : member of the same patent family, corresponding document			

1
EPO FORM 1503 03.82 (P04C01)

**ANNEX TO THE EUROPEAN SEARCH REPORT
ON EUROPEAN PATENT APPLICATION NO.**

EP 13 16 4193

This annex lists the patent family members relating to the patent documents cited in the above-mentioned European search report.
The members are as contained in the European Patent Office EDP file on
The European Patent Office is in no way liable for these particulars which are merely given for the purpose of information.

17-06-2013

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
WO 9407435 A1	14-04-1994	AT 181656 T	15-07-1999
		AU 5127393 A	26-04-1994
		CA 2145990 A1	14-04-1994
		DE 69325518 D1	05-08-1999
		DE 69325518 T2	28-10-1999
		EP 0662808 A1	19-07-1995
		ES 2132259 T3	16-08-1999
		JP H08501715 A	27-02-1996
		WO 9407435 A1	14-04-1994
US 2006041308 A1	23-02-2006	EP 1788981 A2	30-05-2007
		US 2006041308 A1	23-02-2006
		WO 2006023386 A2	02-03-2006
US 2005187623 A1	25-08-2005	EP 1948083 A2	30-07-2008
		US 2005187623 A1	25-08-2005
		US 2007123981 A1	31-05-2007
		WO 2007012978 A2	01-02-2007
US 2004148022 A1	29-07-2004	NONE	

REFERENCES CITED IN THE DESCRIPTION

This list of references cited by the applicant is for the reader's convenience only. It does not form part of the European patent document. Even though great care has been taken in compiling the references, errors or omissions cannot be excluded and the EPO disclaims all liability in this regard.

Patent documents cited in the description

- US 90694407 P [0001]
- US 5748898 A [0062]
- US 5748352 A [0062]
- US 5459570 A [0062]
- US 6111645 A [0062]
- US 6053613 A [0062]